



**PRODUCT MANUAL
FOR PROPOXUR EMULSIFIABLE CONCENTRATE
ACCORDING TO IS 9665 : 1981**

This Product Manual shall be used as reference material by all Regional/Branch Offices & licensees to ensure coherence of practice and transparency in operation of certification under Scheme-I of Bureau of Indian Standards (Conformity Assessment) Regulations, 2018 for various products. The document may also be used by prospective applicants desirous of obtaining BIS certification licence/certificate.

1.	Product	:	IS 9665 : 1981
	Title	:	Specification for Propoxur Emulsifiable Concentrate
	No. of Amendments	:	03
2.	Sampling Guidelines:		
a)	Raw material	:	Propoxur, technical employed in the formulation of Propoxur Emulsifiable Concentrate shall conform to IS 8496.
b)	Grouping guidelines	:	NA (No varieties for the product mentioned in IS)
c)	Sample Size	:	500 ml in original packaging.
3.	List of Test Equipment	:	Please refer ANNEX – <u>A</u>
4.	Scheme of Inspection and Testing	:	Please refer ANNEX – <u>B</u>
5.	Possible tests in a day :		
	i. Description ii. Cold test iii. Flash Point (Abel) iv. Emulsion stability v. Propoxur content vi. Acidity or Alkalinity		
6.	Scope of the Licence :		
	“Licence is granted to use Standard Mark as per IS 9665 : 1981 with the following scope:		
	Name of the product	Propoxure, 20% Emulsifiable Concentrate	

ANNEX - A

TO PRODUCT MANUAL
FOR PROPOXUR EMULSIFIABLE CONCENTRATE
ACCORDING TO IS 9665 : 1981

LIST OF TEST EQUIPMENT

Major test equipment required to test as per the Indian Standard

Sl. No.	Tests used in with Clause Reference	Test Equipment
1.	Description Cl 2.2.1	Glassware, Tap water
2.	Cold Test Cl 2.2.2 (cl 13. 1 of IS 6940)	100 ml glass container with cork/stopper fitted with thermometer, Thermometer Range – 10 to 110 ⁰ C, Refrigerator, Analytical balance with range of 0 to 200gm & Least count of 0.1mg, water bath, Ice-cold water.
3.	Flash Point (Abel) Cl 2.2.3 {IS 1448 (Part 20)}	Cleaning solvent, Coolant, Lubricant, Verification Liquids, Ignitor and pilot light gas, Flash point apparatus/Abel flash point apparatus consisting of test cup, cover assembly, heating vessel, heating device, flash detector, Stirrer, Thermometers 2 (one for the oil cup of range; - 35°C to +70°C, and another for the water bath of the range; -30°C to +80°C), Timing device, Barometer, External cooling bath, Test cup thermal insulating cap, Abel flash point apparatus provided with a stirrer & thermometer, Heating Vessel or bath, Ethylene Glycol.
4.	Emulsion Stability Cl 2.3.4 (cl 13.3 of IS 6940)	Glass Beaker, Capacity 250ml with internal diameter of 6.0 to 6.5 cm and marked at 100 ml, Analytical Balance- Range 0 to 200gms, LC 0.1mg, Mohr-type pipette, 2ml/ 5 ml capacity/ Dropping funnel, Measuring Cylinder, graduated, Capacity 0 to 100ml , Least count 0.5ml, Stop watch 0 to 60 minutes, least count 1sec, Glass Rod, Water Bath with thermometer or digital temp indicator to maintain at 30 ± 1 ⁰ C, Beaker (250 ml), Standard Hard Water, Air conditioner, Hot plate.

<p>5.</p>	<p>Propoxur Content CI 2.3.1 (Appendix A-1 of IS 8496)</p>	<p>Ultra-violet (UV) Spectrophotometric method: UV Spectrophotometer , Volumetric Flasks-100 ml capacity, Pipette- 10 ml, Cuvettes-1 cm quartz, Analytical balance (0 – 200 gm, LC-0.01 mg), Methyl Alcohol, Sodium Methoxide Solutions- 1 N, Sulphuric Acid Solution - 1 N, Propoxur of known percentage of 99 % plus purity. SAPONIFICATION METHOD Apparatus as per Fig .1 of IS 8496, Potassium Hydroxide in Diethylene Glycol - 2 N, Boric Acid Solution- 2 percent (m/v), Nitrogen, Bromocresol green Indicator solution -1 percent (m/v), hydrochloric Acid- 0.1 N, Agate mortar Analytical balance (0- 200 gm, LC-0.01 mg)</p>
<p>6.</p>	<p>Acidity or Alkalinity CI 2.3.2 (CI 13.5 of IS 6940)</p>	<p>Methyl red indicator solution-aqueous – 1 % (m/v), Bromocresol purple indicator solution, Standard Sodium Hydroxide Solution – 0.5 N, Standard Hydrochloric acid, Acetone, Distilled water, Analytical Balance Range 0 to 200gms LC 0.1mg, Hot plate, Range ambient to 100°C/ Heating mantle/Water bath, Whatman filter paper, Conical Flask, 250ml Capacity, Graduated Cylinder, Range 0 to 100ml LC- 1ml, Test Tube, Litmus paper</p>

The above list is indicative only and may not be treated as exhaustive.

ANNEX - B

**SCHEME OF INSPECTION AND TESTING
FOR PROPOXUR EMULSIFIABLE CONCENTRATE
ACCORDING TO IS 9665 : 1981**

1. LABORATORY - A laboratory shall be maintained which shall be suitably equipped (as per the requirement given in column 2 of Table 1) and staffed, where different tests given in the specification shall be carried out in accordance with the methods given in the specification.

1.1 The manufacturer shall prepare a calibration plan for the test equipments.

2. TEST RECORDS – The manufacturer shall maintain test records for the tests carried out to establish conformity.

3. PACKAGING AND MARKING – The Standard Mark, as given in the Schedule of the licence, shall be stenciled/printed on each container of Propoxur EC or printed on the label applied to it, as the case may be, provided always that the material in each container to which this mark is thus applied, conform to every requirement of the specification.

3.1 Packing and marking shall be done as per the provision of the Indian Standard. In addition, the following details shall be mentioned on each container legibly and indelibly:

a) BIS Licence No. CM/L _____.

b) BIS website details i.e – “For details of BIS certification please visit www.bis.gov.in”.

4. CONTROL UNIT – For the purpose of this scheme, the entire quantity of the material processed in a mixer in one operation shall constitute a control unit.

5. LEVELS OF CONTROL - The tests as indicated in column 1 of Table 1 and the levels of control in column 3 of Table 1, shall be carried out on the whole production of the factory which is covered by this plan and appropriate records maintained in accordance with paragraph 2 above.

5.1 All the production which conforms to the Indian Standards and covered by the licence should be marked with Standard Mark.

5.2 On the basis of tests and analysis reports, the decision regarding conformity or otherwise of a control unit to a given requirement shall be made as follows:

5.2.1 A sample shall be drawn from each control unit and tested for all the requirements of the specification. If the sample fails in any of the requirements tested, other than Propoxur Content and/or emulsion stability, the entire control unit represented by the sample shall be considered unfit for the purpose of marking.

5.2.2 In case the sample drawn from the control unit fails in the requirement of Propoxur content and/or emulsion stability but passes in other requirements, the entire quantity of the material in the control unit may be suitably reprocessed and the defects rectified. Such reprocessed material when tested again, shall conform to all the requirements of the specification before it is marked.

6. Raw Materials – The Propoxur technical, used in the manufacture of Propoxur EC shall conform to IS 8496. A test certificate to that effect shall be obtained from the supplier for each consignment of Propoxur, Technical received. Alternatively, a sample from each consignment of the material received shall be tested for its conformity to IS 8496 or ISI marked Propoxur technical may be used.

7. REJECTIONS – Disposal of non-conforming product shall be done in such a way so as to ensure that there is no violation of provisions of BIS Act, 2016.

**TABLE 1
LEVELS OF CONTROL**

(1)				(2)	(3)		
Test Details				Test equipment requirement R: required (or) S: Sub-contracting permitted	Levels of Control		
Cl.	Requirement	Test Method Cl. Ref.	Test Method IS		No. of Sample	Frequency	Remarks
2.2.1	Description	2.2.1	IS 9665	R	One	Every Control Unit	See clause 5.2.1 of SIT
2.2.2	Cold Test	13.1	IS 6940	R	-do-	-do-	
2.2.3	Flash Point (Abel)	-	IS 1448 (Part 20)	R	-do-	-do-	
2.2.4	Emulsion Stability	13.3	IS 6940	R	-do-	-do-	See clause 5.2.2 of SIT
2.3.1	Propoxur Content	Appendix A	IS 8496	R	-do-	-do-	See clause 5.2.1 of SIT
2.3.2	Acidity/Alkalinity	13.5	IS 6940	R	-do-	-do-	See clause 5.2.1 of SIT

Note-1: Whether test equipment is required or sub-contracting is permitted in column 2 shall be decided by the Bureau and shall be mandatory. Subcontracting is permitted to a laboratory recognized by the Bureau or Government laboratories empanelled by the Bureau.

Note-2: Levels of control given in column 3 are only recommendatory in nature. The manufacturer may define the control unit/batch/lot and submit his own levels of control in column 3 with proper justification for approval by BO Head.